

EC DECLARATION OF CONFORMITY

Moor Instruments Ltd
Millwey
Axminster
Devon
EX13 5HU
England

declares that the medical device described hereafter

moorFLPI Full-Field Laser Perfusion Imager

is in conformity with the essential requirements and provisions of Council Directive 93/42/EEC

and is in conformity with the national standard transposing harmonised standards

EN 60601-1: 1990
EN 60601-1-2: 2001
EN 60825-1: 2001

is subject to the procedure set out in Annex 2 of Directive 93/42/EEC under the supervision of Notified Body Number 0120, SGS United Kingdom Limited, Unit 202B, Worle Parkway, Weston-Super-Mare, Somerset BS22 6WA

Axminster, England 27/09/2006



D. Boggett
Managing Director
on behalf of Moor Instruments Ltd