

EC DECLARATION OF CONFORMITY

Moor Instruments Ltd
Millwey
Axminster
Devon
EX13 5HU
England

Declares that the medical device described hereafter

FloLAB/moorLAB Laser Doppler Perfusion Monitor,

is in conformity with the essential requirements and provisions of Council Directive 93/42/EEC and is in conformity with the national standard transposing harmonised standards.

EN 60601-1: 1990
EN 60825-1: 1994

is subject to the procedure set out in Annex 2 of Directive 93/42/EEC under the supervision of Notified Body Number 0120, SGS United Kingdom Limited, Unit 202B, Worle Parkway, Weston-Super-Mare, Somerset BS22 6WA.

Axminster, England 29/05/1998

Managing Director
on behalf of Moor Instruments Ltd